

HORIBA ABX SAS
HORIBA ABX SAS
Parc Euromédecine
Rue du Caducée - BP 7290
34184 MONTPELLIER Cedex 4
FRANCE
Tél.: 33 (0) 4 67 14 15 16
Fax.: 33 (0) 4 67 14 15 17

LOT PX 413
Rev 2

CONTROL

(Exp.) 2018-11-05
(YYYY-MM-DD)

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lysebio | | | | | | | | | | | | | | | |
|--------------------------|---|-------------|-------|--|--|------------|---------|-------|--|------------|---------|---------|-------|------------|-----|---------|-------------------------|
| | | CONTROL | | | | L | CONTROL | | | | N | CONTROL | | | | H | TOLERANCES TOLERANCE |
| | | PENTRA | | | | | PENTRA | | | | | PENTRA | | | | | |
| | | 80 XL80 | XL8 | | | 80 XL80 | XL8 | | | 80 XL80 | XL8 | | | 80 XL80 | XL8 | | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 2.5 | 2.5 | | | ± 0.4 | 7.2 | 7.2 | | | ± 1.0 | 17.5 | 17.5 | | | ± 2.2 | |
| GR RBC | 10 ¹² /mm ³ ; 10 ¹² /l | 2.36 | 2.36 | | | ± 0.16 | 4.64 | 4.64 | | | ± 0.20 | 5.18 | 5.18 | | | ± 0.25 | |
| HB HGB | g/dl | 6.9 | 6.9 | | | ± 0.4 | 14.1 | 14.1 | | | ± 0.5 | 16.6 | 16.6 | | | ± 0.6 | |
| | g/l | 69 | 69 | | | ± 4 | 141 | 141 | | | ± 5 | 166 | 166 | | | ± 6 | |
| | mmol/l | 4.28 | 4.28 | | | ± 0.25 | 8.76 | 8.76 | | | ± 0.31 | 10.31 | 10.31 | | | ± 0.37 | |
| HT HCT | % | 20.1 | 20.1 | | | ± 1.5 | 39.0 | 39.0 | | | ± 2.0 | 45.1 | 46.1 | | | ± 2.5 | |
| | l/l | 0.201 | 0.201 | | | ± 0.015 | 0.390 | 0.390 | | | ± 0.020 | 0.451 | 0.461 | | | ± 0.025 | |
| VGM MCV | µm ³ ; fl | 85 | 85 | | | ± 5 | 84 | 84 | | | ± 5 | 87 | 89 | | | ± 5 | |
| TGMH MCH | pg | 29.2 | 29.2 | | | ± 2.0 | 30.4 | 30.4 | | | ± 2.0 | 32.0 | 32.0 | | | ± 2.5 | |
| | fmol | 1.82 | 1.82 | | | ± 0.12 | 1.89 | 1.89 | | | ± 0.12 | 1.99 | 1.99 | | | ± 0.16 | |
| CCMH MCHC | g/dl | 34.4 | 34.4 | | | ± 3.0 | 36.2 | 36.2 | | | ± 3.0 | 36.8 | 36.0 | | | ± 3.0 | |
| | g/l | 344 | 344 | | | ± 30 | 362 | 362 | | | ± 30 | 368 | 360 | | | ± 30 | |
| | mmol/l | 21.36 | 21.36 | | | ± 1.86 | 22.47 | 22.47 | | | ± 1.86 | 22.87 | 22.36 | | | ± 1.86 | |
| IDR RDW | % | 13.7 | 12.7 | | | ± 4.0 | 13.5 | 12.6 | | | ± 4.0 | 13.8 | 13.0 | | | ± 4.0 | |
| PLAQ. PLTS | 10 ⁹ /mm ³ ; 10 ⁹ /l | 73 | 73 | | | ± 20 | 250 | 250 | | | ± 30 | 495 | 495 | | | ± 50 | |
| VPM MPV | µm ³ ; fl | 9.2 | 9.2 | | | ± 2.0 | 8.9 | 9.1 | | | ± 2.0 | 8.7 | 9.0 | | | ± 2.0 | |
| NEUT | # | 1.50 | 1.50 | | | ± 0.35 | 4.18 | 4.18 | | | ± 0.90 | 12.20 | 12.20 | | | ± 1.90 | |
| | % | 60.1 | 60.1 | | | ± 10.0 | 58.0 | 58.0 | | | ± 10.0 | 69.7 | 69.7 | | | ± 10.0 | |
| LYMPHO | # | 0.66 | 0.66 | | | ± 0.33 | 2.20 | 2.20 | | | ± 0.70 | 2.82 | 2.82 | | | ± 1.50 | |
| | % | 26.4 | 26.4 | | | ± 12.0 | 30.5 | 30.5 | | | ± 8.0 | 16.1 | 16.1 | | | ± 8.0 | |
| MONO | # | 0.08 | 0.08 | | | ± 0.08 | 0.24 | 0.24 | | | ± 0.24 | 0.67 | 0.67 | | | ± 0.67 | |
| | % | 3.3 | 3.3 | | | ± 3.3 | 3.4 | 3.4 | | | ± 3.4 | 3.8 | 3.8 | | | ± 3.8 | |
| EOS | # | 0.17 | 0.17 | | | ± 0.17 | 0.30 | 0.30 | | | ± 0.30 | 1.05 | 1.05 | | | ± 1.05 | |
| | % | 6.7 | 6.7 | | | ± 6.7 | 4.2 | 4.2 | | | ± 4.2 | 6.0 | 6.0 | | | ± 6.0 | |
| BASO | # | 0.09 | 0.09 | | | ± 0.09 | 0.28 | 0.28 | | | ± 0.28 | 0.77 | 0.77 | | | ± 0.77 | |
| | % | 3.5 | 3.5 | | | ± 3.5 | 3.9 | 3.9 | | | ± 3.9 | 4.4 | 4.4 | | | ± 4.4 | |

LOT PX 413
Rev 2

CONTROL

(Exp.) 2018-11-05
(YYYY-MM-DD)

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lysebio | | | | | | | | | | | | | | | |
|--------------------------|---|--------------------|-------|-------|----|--------------------|---------|-------|-------|--------------------|------|---------|-------|--------------------|-------|---------|-------------------------|
| | | CONTROL | | | | L | CONTROL | | | | N | CONTROL | | | | H | TOLERANCES TOLERANCE |
| | | PENTRA | | | | | PENTRA | | | | | PENTRA | | | | | |
| | | 60 60C+ ES60 | MS60 | MSCR | CP | 60 60C+ ES60 | MS60 | MSCR | CP | 60 60C+ ES60 | MS60 | MSCR | CP | 60 60C+ ES60 | MS60 | MSCR | CP |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 2.4 | 2.4 | 2.5 | | ± 0.4 | 7.2 | 7.2 | 7.3 | | | ± 1.0 | 17.6 | 17.4 | 18.0 | ± 2.2 | |
| GR RBC | 10 ¹² /mm ³ ; 10 ¹² /l | 2.41 | 2.37 | 2.40 | | ± 0.16 | 4.61 | 4.64 | 4.64 | | | ± 0.20 | 5.12 | 5.15 | 5.19 | ± 0.25 | |
| HB HGB | g/dl | 6.9 | 6.8 | 6.9 | | ± 0.4 | 14.0 | 14.0 | 13.9 | | | ± 0.5 | 16.5 | 16.5 | 16.5 | ± 0.6 | |
| | g/l | 69 | 68 | 69 | | ± 4 | 140 | 140 | 139 | | | ± 5 | 165 | 165 | 165 | ± 6 | |
| | mmol/l | 4.28 | 4.22 | 4.28 | | ± 0.25 | 8.69 | 8.69 | 8.63 | | | ± 0.31 | 10.25 | 10.25 | 10.25 | ± 0.37 | |
| HT HCT | % | 19.8 | 19.4 | 19.4 | | ± 1.5 | 38.7 | 38.5 | 38.3 | | | ± 2.0 | 45.1 | 45.3 | 45.7 | ± 2.5 | |
| | l/l | 0.198 | 0.194 | 0.194 | | ± 0.015 | 0.387 | 0.385 | 0.383 | | | ± 0.020 | 0.451 | 0.453 | 0.457 | ± 0.025 | |
| VGM MCV | µm ³ ; fl | 82 | 82 | 81.0 | | ± 5 | 84 | 83 | 82.5 | | | ± 5 | 88 | 88 | 88.0 | ± 5 | |
| TGMH MCH | pg | 28.6 | 28.7 | 28.8 | | ± 2.0 | 30.4 | 30.2 | 30.0 | | | ± 2.0 | 32.2 | 32.0 | 31.8 | ± 2.5 | |
| | fmol | 1.78 | 1.78 | 1.79 | | ± 0.12 | 1.89 | 1.87 | 1.86 | | | ± 0.12 | 2.00 | 1.99 | 1.97 | ± 0.16 | |
| CCMH MCHC | g/dl | 34.9 | 35.0 | 35.5 | | ± 3.0 | 36.2 | 36.4 | 36.3 | | | ± 3.0 | 36.6 | 36.4 | 36.1 | ± 3.0 | |
| | g/l | 349 | 350 | 355 | | ± 30 | 362 | 364 | 363 | | | ± 30 | 366 | 364 | 361 | ± 30 | |
| | mmol/l | 21.68 | 21.73 | 22.04 | | ± 1.86 | 22.45 | 22.57 | 22.55 | | | ± 1.86 | 22.74 | 22.61 | 22.43 | ± 1.86 | |
| IDR RDW | % | 13.5 | 12.8 | 12.8 | | ± 4.0 | 13.2 | 12.5 | 12.4 | | | ± 4.0 | 13.3 | 12.3 | 12.3 | ± 4.0 | |
| PLAQ. PLTS | 10 ⁹ /mm ³ ; 10 ⁹ /l | 75 | 74 | 72 | | ± 20 | 250 | 255 | 245 | | | ± 30 | 490 | 505 | 480 | ± 50 | |
| VPM MPV | µm ³ ; fl | 9.0 | 9.0 | 8.8 | | ± 2.0 | 8.8 | 8.9 | 8.3 | | | ± 2.0 | 8.5 | 8.7 | 8.1 | ± 2.0 | |
| NEUT | # | 1.44 | 1.46 | 1.53 | | ± 0.35 | 4.17 | 4.19 | 4.34 | | | ± 0.90 | 12.32 | 12.23 | 12.96 | ± 1.90 | |
| | % | 59.8 | 60.9 | 61.0 | | ± 10.0 | 57.9 | 58.2 | 59.5 | | | ± 10.0 | 70.0 | 70.3 | 72.0 | ± 10.0 | |
| LYMPHO | # | 0.65 | 0.62 | 0.66 | | ± 0.33 | 2.24 | 2.18 | 2.23 | | | ± 0.70 | 2.80 | 2.64 | 2.70 | ± 1.50 | |
| | % | 27.1 | 26.0 | 26.5 | | ± 12.0 | 31.1 | 30.3 | 30.5 | | | ± 8.0 | 15.9 | 15.2 | 15.0 | ± 8.0 | |
| MONO | # | 0.07 | 0.07 | 0.08 | | ± 0.07 | 0.22 | 0.25 | 0.22 | | | ± 0.22 | 0.65 | 0.68 | 0.54 | ± 0.54 | |
| | % | 3.0 | 3.1 | 3.0 | | ± 3.0 | 3.1 | 3.5 | 3.0 | | | ± 3.0 | 3.7 | 3.9 | 3.0 | ± 3.0 | |
| EOS | # | 0.16 | 0.16 | 0.15 | | ± 0.15 | 0.30 | 0.32 | 0.26 | | | ± 0.26 | 1.04 | 1.08 | 0.99 | ± 0.99 | |
| | % | 6.6 | 6.5 | 6.0 | | ± 6.0 | 4.1 | 4.4 | 3.5 | | | ± 3.5 | 5.9 | 6.2 | 5.5 | ± 5.5 | |
| BASO | # | 0.08 | 0.08 | 0.09 | | ± 0.08 | 0.27 | 0.26 | 0.26 | | | ± 0.26 | 0.79 | 0.77 | 0.81 | ± 0.77 | |
| | % | 3.5 | 3.5 | 3.5 | | ± 3.5 | 3.8 | 3.6 | 3.5 | | | ± 3.5 | 4.5 | 4.4 | 4.5 | ± 4.4 | |

Ref: TEMP-0821 Rev.42 FRONT / RECTO 1300040596