


**LOT** PX 413

**CONTROL**

 (Exp.) 2018-11-05  
(YYYY-MM-DD)

Rev 2

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio																	
		CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA				120 RETIC		PENTRA				120 RETIC		PENTRA				120 RETIC	
		120	DX120	DX NEXUS	DF NEXUS		120	DX120	DX NEXUS	DF NEXUS	120		DX120	DX NEXUS	DF NEXUS				
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.4	2.4	2.4		± 0.4	7.4	7.4	7.4		± 1.0	18.6	18.6	18.6		± 2.2			
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.38	2.38	2.38		± 0.16	4.64	4.64	4.64		± 0.20	5.25	5.25	5.25		± 0.25			
HB HGB	g/dl	7.0	7.0	7.0		± 0.4	14.0	14.0	14.0		± 0.5	16.5	16.5	16.5		± 0.6			
	g/l	70	70	70		± 4	140	140	140		± 5	165	165	165		± 6			
	mmol/l	4.35	4.35	4.35		± 0.25	8.69	8.69	8.69		± 0.31	10.25	10.25	10.25		± 0.37			
HT HCT	%	20.0	20.0	20.0		± 1.5	39.0	39.0	39.0		± 2.0	45.7	45.7	45.7		± 2.5			
	l/l	0.200	0.200	0.200		± 0.015	0.390	0.390	0.390		± 0.020	0.457	0.457	0.457		± 0.025			
VGM MCV	µm <sup>3</sup> ; fl	84	84	84		± 5	84	84	84		± 5	87	87	87		± 5			
TGMH MCH	pg	29.4	29.4	29.4		± 2.0	30.2	30.2	30.2		± 2.0	31.4	31.4	31.4		± 2.5			
	fmol	1.83	1.83	1.83		± 0.12	1.87	1.87	1.87		± 0.12	1.95	1.95	1.95		± 0.16			
CCMH MCHC	g/dl	35.0	35.0	35.0		± 3.0	35.9	35.9	35.9		± 3.0	36.1	36.1	36.1		± 3.0			
	g/l	350	350	350		± 30	359	359	359		± 30	361	361	361		± 30			
	mmol/l	21.74	21.74	21.74		± 1.86	22.31	22.31	22.31		± 1.86	22.43	22.43	22.43		± 1.86			
IDR RDW	%	15.5	15.5	15.5		± 4.0	16.0	16.0	16.0		± 4.0	14.8	14.8	14.8		± 4.0			
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	75	75	75		± 20	252	252	252		± 30	490	490	490		± 50			
VPM MPV	µm <sup>3</sup> ; fl	9.0	9.0	9.0		± 2.0	8.8	8.8	8.8		± 2.0	8.8	8.8	8.8		± 2.0			
NEUT	#	1.46	1.49	1.49		± 0.35	4.39	4.40	4.40		± 0.90	13.70	13.70	13.70		± 1.90			
	%	60.8	62.1	62.1		± 10.0	59.3	59.4	59.4		± 10.0	73.6	73.6	73.6		± 10.0			
LYMPHO	#	0.62	0.56	0.56		± 0.33	2.23	2.09	2.09		± 0.70	2.72	2.51	2.51		± 1.50			
	%	25.8	23.4	23.4		± 12.0	30.1	28.2	28.2		± 8.0	14.6	13.5	13.5		± 8.0			
MONO	#	0.10	0.10	0.10		± 0.10	0.29	0.30	0.30		± 0.29	0.86	0.86	0.86		± 0.86			
	%	4.2	4.3	4.3		± 4.2	3.9	4.0	4.0		± 3.9	4.6	4.6	4.6		± 4.6			
EOS	#	0.16	0.17	0.17		± 0.16	0.27	0.30	0.30		± 0.27	0.87	0.89	0.89		± 0.87			
	%	6.7	7.1	7.1		± 6.7	3.7	4.0	4.0		± 3.7	4.7	4.8	4.8		± 4.7			
BASO	#	0.06	0.07	0.07		± 0.06	0.22	0.33	0.33		± 0.22	0.47	0.65	0.65		± 0.47			
	%	2.5	3.1	3.1		± 2.5	3.0	4.4	4.4		± 3.0	2.5	3.5	3.5		± 2.5			

Ref: TEMP-0821 Rev.42 BACK / VERSO 1300040596